Email : sales@scilabexport.com
Phone: +91-7082934803

Product Name :

Disposable Endotracheal Tube Without Cuff Sterile

Product Code :

SLE/RAE/06



Email : sales@scilabexport.com
Phone: +91-7082934803

Description :

Disposable Endotracheal Tube Without Cuff Sterile (Product Code: SLE/RAE/06) is an uncuffed, single-use PVC tube inserted through the mouth or nose into the trachea so that a patient can be ventilated manually or mechanically during general anaesthesia or resuscitation.Its small 3.5 mm internal diameter and plain shaft, with no cuff or inflation line, set it apart fromthe cuffed tube in the same range. It is manufactured by **Jain Scientific Equipments Pvt Ltd (ScienceLab)**, Ambala, India, and supplied to hospitals and health programmes.

Specifications

Parameter	Detail
Material	Polyvinyl chloride (PVC)
Type	Endotracheal tube without cuff; disposable, sterile
Size	3.5 mm internal diameter
Tip	Open distal end, Magill-type point with oral angle of 37.5°
Murphy’s eye	Yes
Markings	Graduations
Connector	Standard connector, 15 mm external diameter, at the proximal end for the ventilation system
Initial sterilisation method	Ethylene oxide gas or gamma radiation
Cuff	Uncuffed
Pack	Individual single-use pack
Product code	SLE/RAE/06
Manufacturer	Jain Scientific Equipments Pvt Ltd (ScienceLab), Ambala, India

What Is the Uncuffed Endotracheal Tube Used For?

Like any endotracheal tube, this one creates a secure passage from the upper airway into the trachea, so breaths can be delivered by a hand resuscitator or a ventilator while a patient is anaesthetised or being resuscitated. Tube choice for adults or children depends on the diameter, and the clinician selects the size and cuff type for each patient.

Without a cuff there is no pilot balloon to inflate or check: the tube is positioned, its depth noted against the graduations, and the 15 mm connector fitted to the breathing system. Lubricate the distal end of the tube, and the distal half of the introducer (bougie) if one is used, before insertion. Insertion must be carried out only by staff trained in resuscitation practice.

What to Specify When Ordering

- The internal diameter – this listing is 3.5 mm; ask us about other sizes.
-
-

Man